

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV103119\
 Data File : VV013457.D
 Acq On : 31 Oct 2019 20:13
 Operator : SY/MD
 Sample : K5635-01
 Misc : 25.0ML/MSVOA V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BE9G6

Quant Time: Nov 01 06:02:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR103119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 01 05:32:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	339870	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	359063	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	189037	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	126916	5.07	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.40%
7) Chloroethane-d5	1.58	69	106554	5.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.20%
11) 1,1-Dichloroethene-d2	2.13	63	169243	3.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.60%
20) 2-Butanone-d5	3.93	46	397623	51.24	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.48%
24) Chloroform-d	4.40	84	286403	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.40%
26) 1,2-Dichloroethane-d4	5.08	65	154430	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.80%
32) Benzene-d6	5.10	84	568023	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
36) 1,2-Dichloropropane-d6	6.12	67	184142	4.94	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.80%
41) Toluene-d8	7.36	98	480555	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.20%
43) trans-1,3-Dichloropropene-	7.66	79	64977	4.35	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.00%
46) 2-Hexanone-d5	8.13	63	229307	41.74	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.48%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	123522	4.32	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	197780	5.15	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.00%

Target Compounds					Ovalue
13) Acetone	2.19	43	39015	11.096 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV103119\
Data File : VV013457.D
Acq On : 31 Oct 2019 20:13
Operator : SY/MD
Sample : K5635-01
Misc : 25.0ML/MSVOA V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BE9G6

Quant Time: Nov 01 06:02:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR103119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 01 05:32:11 2019
Response via : Initial Calibration

