

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV103122\
 Data File : VV028787.D
 Acq On : 31 Oct 2022 17:54
 Operator : SY/MD
 Sample : N5354-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YBMP6

Quant Time: Nov 01 04:57:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 01 04:42:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	432544	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	391167	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	168472	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	132532	32.322	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	64.640%		
7) Chloroethane-d5	1.564	69	135551	38.671	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	77.340%		
11) 1,1-Dichloroethene-d2	2.104	63	215614	27.643	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	55.280%#		
21) 2-Butanone-d5	3.876	46	244611	81.951	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	81.950%		
24) Chloroform-d	4.342	84	249389	40.326	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	80.660%		
26) 1,2-Dichloroethane-d4	5.027	65	178162	43.560	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	87.120%		
32) Benzene-d6	5.046	84	515752	42.236	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	84.480%		
36) 1,2-Dichloropropane-d6	6.063	67	183968	43.882	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	87.760%		
41) Toluene-d8	7.310	98	440532	41.036	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	82.080%		
43) trans-1,3-Dichloroprop...	7.619	79	63230	39.269	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	78.540%		
47) 2-Hexanone-d5	8.085	63	154757	81.685	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	81.690%		
56) 1,1,2,2-Tetrachloroeth...	10.207	84	235537	42.606	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	85.220%		
66) 1,2-Dichlorobenzene-d4	11.615	152	146728	44.122	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	88.240%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.114	96	4488	1.391	ug/L #	1
13) Acetone	2.166	43	192927	66.757	ug/L	99
34) Trichloroethene	5.911	95	99613	30.931	ug/L	99
46) Tetrachloroethene	7.972	164	9009	4.036	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV103122\
Data File : VV028787.D
Acq On : 31 Oct 2022 17:54
Operator : SY/MD
Sample : N5354-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YBMP6

Quant Time: Nov 01 04:57:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 01 04:42:28 2022
Response via : Initial Calibration

