

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV103122\
 Data File : VV028797.D
 Acq On : 31 Oct 2022 21:56
 Operator : SY/MD
 Sample : N5321-12
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBWN6

Quant Time: Nov 01 04:58:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 01 04:42:28 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	415594	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	377214	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.243	152	160872	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	133226	33.816	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	67.640%	
7) Chloroethane-d5	1.564	69	127762	37.935	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	75.880%	
11) 1,1-Dichloroethene-d2	2.105	63	217186	28.980	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	57.960%#	
21) 2-Butanone-d5	3.873	46	262644	91.581	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	91.580%	
24) Chloroform-d	4.343	84	247645	41.677	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	83.360%	
26) 1,2-Dichloroethane-d4	5.027	65	179397	45.651	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.300%	
32) Benzene-d6	5.047	84	515343	43.763	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.520%	
36) 1,2-Dichloropropane-d6	6.066	67	184648	45.673	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	91.340%	
41) Toluene-d8	7.310	98	436043	42.121	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	84.240%	
43) trans-1,3-Dichloroprop...	7.619	79	63086	40.629	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	81.260%	
47) 2-Hexanone-d5	8.085	63	151825	83.102	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	83.100%	
56) 1,1,2,2-Tetrachloroeth...	10.207	84	235923	44.255	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	88.500%	
66) 1,2-Dichlorobenzene-d4	11.616	152	143158	45.082	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	90.160%	
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.310	62	43665	9.345	ug/L	96
12) 1,1-Dichloroethene	2.118	96	8882	2.865	ug/L #	1
13) Acetone	2.163	43	6898	2.484	ug/L	93
17) trans-1,2-Dichloroethene	2.757	96	216726	72.284	ug/L	98
20) cis-1,2-Dichloroethene	3.902	96	762098	230.368	ug/L #	98
34) Trichloroethene	5.908	95	325798	104.907	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV103122\
Data File : VV028797.D
Acq On : 31 Oct 2022 21:56
Operator : SY/MD
Sample : N5321-12
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBWN6

Quant Time: Nov 01 04:58:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 01 04:42:28 2022
Response via : Initial Calibration

