

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV103123\
 Data File : VV032797.D
 Acq On : 31 Oct 2023 14:41
 Operator : SY/MD
 Sample : 05105-11
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0005

Quant Time: Nov 01 05:48:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 01 05:44:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	294518	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	326438	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	145049	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	149822	53.484	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 106.960%		
7) Chloroethane-d5	1.532	69	127415	54.385	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 108.760%		
11) 1,1-Dichloroethene-d2	2.060	65	58930	50.932	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 101.860%		
21) 2-Butanone-d5	3.786	46	137531	113.189	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 113.190%		
24) Chloroform-d	4.249	84	234227	52.385	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 104.760%		
26) 1,2-Dichloroethane-d4	4.944	65	150719	54.236	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 108.480%		
32) Benzene-d6	4.963	84	467836	55.740	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 111.480%		
36) 1,2-Dichloropropane-d6	5.989	67	143836	56.408	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 112.820%		
41) Toluene-d8	7.246	98	447711	54.173	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 108.340%		
43) trans-1,3-Dichloroprop...	7.555	79	70301	52.937	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 105.880%		
47) 2-Hexanone-d5	8.027	63	102189	109.774	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 109.770%		
56) 1,1,2,2-Tetrachloroeth...	10.156	84	203062	52.478	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 104.960%		
66) 1,2-Dichlorobenzene-d4	11.564	152	170224	60.899	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 121.800%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV103123\
Data File : VV032797.D
Acq On : 31 Oct 2023 14:41
Operator : SY/MD
Sample : 05105-11
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0005

Quant Time: Nov 01 05:48:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 01 05:44:49 2023
Response via : Initial Calibration

