

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV103123\
 Data File : VV032799.D
 Acq On : 31 Oct 2023 15:27
 Operator : SY/MD
 Sample : 05105-08
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0010

Quant Time: Nov 01 05:49:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 01 05:44:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	332314	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	368127	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	164437	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	145821	46.135	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.280%
7) Chloroethane-d5	1.532	69	118653	44.885	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.760%
11) 1,1-Dichloroethene-d2	2.060	65	58430	44.756	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.520%
21) 2-Butanone-d5	3.786	46	125434	91.492	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.490%
24) Chloroform-d	4.253	84	229706	45.531	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.060%
26) 1,2-Dichloroethane-d4	4.944	65	148925	47.495	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.000%
32) Benzene-d6	4.960	84	448198	47.353	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.700%
36) 1,2-Dichloropropane-d6	5.992	67	136304	47.401	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.800%
41) Toluene-d8	7.246	98	430224	46.162	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.320%
43) trans-1,3-Dichloroprop...	7.555	79	67928	45.357	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.720%
47) 2-Hexanone-d5	8.027	63	92655	88.261	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.260%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	199027	45.611	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.220%
66) 1,2-Dichlorobenzene-d4	11.564	152	165052	52.086	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.180%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV103123\
Data File : VV032799.D
Acq On : 31 Oct 2023 15:27
Operator : SY/MD
Sample : 05105-08
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0010

Quant Time: Nov 01 05:49:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 01 05:44:49 2023
Response via : Initial Calibration

