

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW103123\
 Data File : VW032801.D
 Acq On : 31 Oct 2023 16:14
 Operator : SY/MD
 Sample : 05105-10
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0004

Quant Time: Nov 01 05:49:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 01 05:44:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	340843	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	376780	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	203961	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	143339	44.215	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.440%
7) Chloroethane-d5	1.532	69	120971	44.617	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.240%
11) 1,1-Dichloroethene-d2	2.060	65	59579	44.494	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.980%
21) 2-Butanone-d5	3.790	46	127978	91.012	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.010%
24) Chloroform-d	4.253	84	233388	45.103	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.200%
26) 1,2-Dichloroethane-d4	4.944	65	146523	45.560	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.120%
32) Benzene-d6	4.963	84	451865	46.644	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.280%
36) 1,2-Dichloropropane-d6	5.992	67	137022	46.556	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.120%
41) Toluene-d8	7.246	98	442262	46.364	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.720%
43) trans-1,3-Dichloroprop...	7.555	79	67652	44.136	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.280%
47) 2-Hexanone-d5	8.027	63	100340	93.386	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.390%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	207157	46.384	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.760%
66) 1,2-Dichlorobenzene-d4	11.564	152	199767	50.825	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.660%
Target Compounds						
					Qvalue	
52) Ethylbenzene	8.953	91	40436	3.222	ug/L	99
60) Isopropylbenzene	9.870	105	67892	5.489	ug/L	98
63) 1,2,4-Trimethylbenzene	10.854	105	1986883	213.394	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV103123\
Data File : VV032801.D
Acq On : 31 Oct 2023 16:14
Operator : SY/MD
Sample : 05105-10
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0004

Quant Time: Nov 01 05:49:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 01 05:44:49 2023
Response via : Initial Calibration

