

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV103123\
 Data File : VV032803.D
 Acq On : 31 Oct 2023 17:01
 Operator : SY/MD
 Sample : 05132-16ME
 Misc : 5.48g/5.0mL/100ul/5.0mL/MSVOA_V/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH5Q2ME

Quant Time: Nov 01 05:50:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 01 05:44:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	449481	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	436557	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	214559	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	140277	32.812	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	65.620%
7) Chloroethane-d5	1.507	69	82622	23.108	ug/L	-0.03
Spiked Amount	50.000	Range	70 - 130	Recovery	=	46.220%#
11) 1,1-Dichloroethene-d2	2.037	65	69303	39.247	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.500%
21) 2-Butanone-d5	3.805	46	130632	70.446	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	70.450%
24) Chloroform-d	4.249	84	259746	38.064	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	76.120%
26) 1,2-Dichloroethane-d4	4.940	65	167402	39.471	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.940%
32) Benzene-d6	4.957	84	516240	45.992	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.980%
36) 1,2-Dichloropropane-d6	5.989	67	153176	44.919	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.840%
41) Toluene-d8	7.243	98	450787	40.787	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.580%
43) trans-1,3-Dichloroprop...	7.555	79	72322	40.722	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.440%
47) 2-Hexanone-d5	8.046	63	102574	82.394	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.390%
56) 1,1,2,2-Tetrachloroeth...	10.159	84	203885	39.400	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	78.800%
66) 1,2-Dichlorobenzene-d4	11.567	152	179473	43.407	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV103123\
Data File : VV032803.D
Acq On : 31 Oct 2023 17:01
Operator : SY/MD
Sample : 05132-16ME
Misc : 5.48g/5.0mL/100ul/5.0mL/MSVOA_V/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH5Q2ME

Quant Time: Nov 01 05:50:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 01 05:44:49 2023
Response via : Initial Calibration

