

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV110118\
 Data File : VV008506.D
 Acq On : 01 Nov 2018 12:00
 Operator : SY/MD
 Sample : J5669-12
 Misc : 5.0 mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Nov 02 07:27:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM103118WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 02 07:10:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	209983	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	186280	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	80445	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	57677	43.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.40%
7) Chloroethane-d5	1.56	69	38581	51.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	103.80%
11) 1,1-Dichloroethene-d2	2.13	63	74910	34.92	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.84%
21) 2-Butanone-d5	3.94	46	119983	92.83	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.83%
24) Chloroform-d	4.41	84	130660	44.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.18%
26) 1,2-Dichloroethane-d4	5.09	65	90872	46.41	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.82%
32) Benzene-d6	5.10	84	269275	47.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.48%
36) 1,2-Dichloropropane-d6	6.12	67	90545	47.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.42%
41) Toluene-d8	7.36	98	241002	46.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.54%
43) trans-1,3-Dichloropropene-	7.67	79	43714	45.41	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.82%
47) 2-Hexanone-d5	8.14	63	88946	92.96	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.96%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	123138	46.90	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.80%
64) 1,2-Dichlorobenzene-d4	11.68	152	80845	45.84	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.68%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV110118\
Data File : VV008506.D
Acq On : 01 Nov 2018 12:00
Operator : SY/MD
Sample : J5669-12
Misc : 5.0 mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Nov 02 07:27:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM103118WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 02 07:10:33 2018
Response via : Initial Calibration

