

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV110118\
 Data File : VV008507.D
 Acq On : 01 Nov 2018 12:24
 Operator : SY/MD
 Sample : J5716-13
 Misc : 5.0 mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 ETBZ0

Quant Time: Nov 02 07:31:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM103118WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 02 07:10:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	208915	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	184968	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	78170	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	56722	42.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.40%
7) Chloroethane-d5	1.56	69	40316	54.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.02%
11) 1,1-Dichloroethene-d2	2.13	63	73483	34.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.86%
21) 2-Butanone-d5	3.94	46	115386	89.73	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.73%
24) Chloroform-d	4.41	84	129382	44.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.76%
26) 1,2-Dichloroethane-d4	5.09	65	89209	45.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.58%
32) Benzene-d6	5.10	84	264685	47.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.52%
36) 1,2-Dichloropropane-d6	6.12	67	90298	47.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.82%
41) Toluene-d8	7.36	98	235183	45.97	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.94%
43) trans-1,3-Dichloropropene-	7.67	79	42269	44.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.44%
47) 2-Hexanone-d5	8.13	63	86274	90.81	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.81%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	119309	45.76	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.52%
64) 1,2-Dichlorobenzene-d4	11.68	152	79608	46.46	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.92%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV110118\
Data File : VV008507.D
Acq On : 01 Nov 2018 12:24
Operator : SY/MD
Sample : J5716-13
Misc : 5.0 mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
ETBZ0

Quant Time: Nov 02 07:31:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM103118WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 02 07:10:33 2018
Response via : Initial Calibration

