

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV110118\
 Data File : VV008503.D
 Acq On : 01 Nov 2018 10:38
 Operator : SY/MD
 Sample : VV1101WBL01
 Misc : 5.0 mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK79

Quant Time: Nov 02 07:17:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM103118WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 02 07:10:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	224379	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	199190	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	80184	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	57579	40.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.72%
7) Chloroethane-d5	1.58	69	47565	59.88	ug/L	0.01
Spiked Amount	50.000	Range	70 - 130	Recovery	=	119.76%
11) 1,1-Dichloroethene-d2	2.13	63	81653	35.62	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.24%
21) 2-Butanone-d5	3.93	46	128039	92.71	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.71%
24) Chloroform-d	4.41	84	144074	46.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.02%
26) 1,2-Dichloroethane-d4	5.08	65	99288	47.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.90%
32) Benzene-d6	5.10	84	293708	48.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.38%
36) 1,2-Dichloropropane-d6	6.12	67	99069	48.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.64%
41) Toluene-d8	7.36	98	262235	47.60	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.20%
43) trans-1,3-Dichloropropene-	7.67	79	48198	46.82	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.64%
47) 2-Hexanone-d5	8.13	63	87594	85.62	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.62%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	125398	44.67	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.34%
64) 1,2-Dichlorobenzene-d4	11.68	152	83984	47.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.56%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV110118\
Data File : VV008503.D
Acq On : 01 Nov 2018 10:38
Operator : SY/MD
Sample : VV1101WBL01
Misc : 5.0 mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK79

Quant Time: Nov 02 07:17:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM103118WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 02 07:10:33 2018
Response via : Initial Calibration

