

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV110118\
 Data File : VV008504.D
 Acq On : 01 Nov 2018 11:14
 Operator : SY/MD
 Sample : J5668-21
 Misc : 5.0 mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Nov 02 07:20:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM103118WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 02 07:10:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	215109	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	190989	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	81113	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	58375	42.68	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.36%
7) Chloroethane-d5	1.57	69	27488	36.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	72.20%
11) 1,1-Dichloroethene-d2	2.13	63	76492	34.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.62%
21) 2-Butanone-d5	3.93	46	112334	84.84	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	84.84%
24) Chloroform-d	4.40	84	131117	43.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.36%
26) 1,2-Dichloroethane-d4	5.08	65	91568	45.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.30%
32) Benzene-d6	5.10	84	269732	46.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.28%
36) 1,2-Dichloropropane-d6	6.12	67	90758	46.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.28%
41) Toluene-d8	7.36	98	240176	45.47	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.94%
43) trans-1,3-Dichloropropene-	7.67	79	43274	43.84	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.68%
47) 2-Hexanone-d5	8.13	63	83334	84.95	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.95%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	117121	43.51	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	87.02%
64) 1,2-Dichlorobenzene-d4	11.68	152	79905	44.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.88%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV110118\
Data File : VV008504.D
Acq On : 01 Nov 2018 11:14
Operator : SY/MD
Sample : J5668-21
Misc : 5.0 mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Nov 02 07:20:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM103118WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 02 07:10:33 2018
Response via : Initial Calibration

