

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110121\
 Data File : VV023108.D
 Acq On : 01 Nov 2021 11:35
 Operator : SY/MD
 Sample : VV1101WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK237

Quant Time: Nov 02 01:14:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 02 01:11:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	134094	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	132146	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	63445	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	35898	3.064	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	61.200%
7) Chloroethane-d5	1.564	69	34686	4.786	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.800%
11) 1,1-Dichloroethene-d2	2.104	63	56391	3.336	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.800%
20) 2-Butanone-d5	3.915	46	110787	58.941	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.880%
24) Chloroform-d	4.346	84	98632	5.178	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.600%
26) 1,2-Dichloroethane-d4	5.030	65	48873	5.443	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.800%
32) Benzene-d6	5.047	84	191943	4.975	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.600%
36) 1,2-Dichloropropane-d6	6.069	67	58785	4.950	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.000%
41) Toluene-d8	7.313	98	162904	4.701	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.000%
43) trans-1,3-Dichloroprop...	7.622	79	22526	5.413	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	108.200%
46) 2-Hexanone-d5	8.088	63	98342	63.833	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	127.660%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	44983	5.485	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	64241	5.675	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110121\
Data File : VV023108.D
Acq On : 01 Nov 2021 11:35
Operator : SY/MD
Sample : VV1101WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK237

Quant Time: Nov 02 01:14:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 02 01:11:36 2021
Response via : Initial Calibration

