

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW110121\
 Data File : VW023118.D
 Acq On : 01 Nov 2021 15:44
 Operator : SY/MD
 Sample : M4400-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG333

Quant Time: Nov 02 01:16:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 02 01:11:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	125501	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	129991	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62767	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	28908	2.636	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	52.800%	
7) Chloroethane-d5	1.567	69	28571	4.212	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	84.200%	
11) 1,1-Dichloroethene-d2	2.108	63	47791	3.021	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	60.400%	
20) 2-Butanone-d5	3.889	46	104302	59.290	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	118.580%	
24) Chloroform-d	4.352	84	87310	4.897	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.000%	
26) 1,2-Dichloroethane-d4	5.034	65	42542	5.062	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	101.200%	
32) Benzene-d6	5.053	84	162307	4.277	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	85.600%	
36) 1,2-Dichloropropane-d6	6.072	67	51381	4.398	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	88.000%	
41) Toluene-d8	7.316	98	138122	4.052	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	81.000%	
43) trans-1,3-Dichloroprop...	7.625	79	19677	4.807	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	96.200%	
46) 2-Hexanone-d5	8.088	63	84649	55.856	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	111.720%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	39125	4.850	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	97.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	56853	5.077	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	101.600%	
Target Compounds						Qvalue
13) Acetone	2.175	43	2943	3.410	ug/L	72
42) Toluene	7.397	91	6010	0.170	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110121\
Data File : VV023118.D
Acq On : 01 Nov 2021 15:44
Operator : SY/MD
Sample : M4400-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG333

Quant Time: Nov 02 01:16:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 02 01:11:36 2021
Response via : Initial Calibration

