

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110123\
 Data File : VV032821.D
 Acq On : 01 Nov 2023 11:42
 Operator : SY/MD
 Sample : 05105-18
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Nov 02 01:10:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Nov 02 01:08:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	342829	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	376385	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	168780	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	139528	42.790	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.580%
7) Chloroethane-d5	1.532	69	120238	44.089	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.180%
11) 1,1-Dichloroethene-d2	2.060	65	58237	43.240	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.480%
21) 2-Butanone-d5	3.786	46	128717	91.007	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.010%
24) Chloroform-d	4.252	84	230695	44.324	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.640%
26) 1,2-Dichloroethane-d4	4.944	65	146966	45.433	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.860%
32) Benzene-d6	4.963	84	456537	47.175	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.360%
36) 1,2-Dichloropropane-d6	5.989	67	137826	46.879	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.760%
41) Toluene-d8	7.246	98	428927	45.013	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.020%
43) trans-1,3-Dichloroprop...	7.554	79	70604	46.110	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.220%
47) 2-Hexanone-d5	8.027	63	93335	86.958	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.960%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	200928	45.036	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.080%
66) 1,2-Dichlorobenzene-d4	11.564	152	166178	51.092	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.180%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110123\
Data File : VV032821.D
Acq On : 01 Nov 2023 11:42
Operator : SY/MD
Sample : 05105-18
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Nov 02 01:10:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Nov 02 01:08:52 2023
Response via : Initial Calibration

