

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110221\
 Data File : VV023176.D
 Acq On : 03 Nov 2021 10:40
 Operator : SY/MD
 Sample : VV1102WBL03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 72 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK240

Quant Time: Nov 07 01:02:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 02 23:31:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	117957	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	118418	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	55934	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	43979	4.267	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.400%
7) Chloroethane-d5	1.564	69	36774	5.768	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.400%
11) 1,1-Dichloroethene-d2	2.105	63	63230	4.252	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.000%
20) 2-Butanone-d5	3.905	46	73218	44.282	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.560%
24) Chloroform-d	4.346	84	90095	5.377	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.600%
26) 1,2-Dichloroethane-d4	5.031	65	42411	5.369	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.400%
32) Benzene-d6	5.047	84	175944	5.089	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.800%
36) 1,2-Dichloropropane-d6	6.066	67	53860	5.061	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.200%
41) Toluene-d8	7.313	98	150469	4.845	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.000%
43) trans-1,3-Dichloroprop...	7.622	79	18506	4.963	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.200%
46) 2-Hexanone-d5	8.088	63	66489	48.160	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.320%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	34361	4.676	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.600%
66) 1,2-Dichlorobenzene-d4	11.625	152	56872	5.699	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.000%
Target Compounds						Qvalue
16) Methylene chloride	2.506	84	3570	0.538	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110221\
Data File : VV023176.D
Acq On : 03 Nov 2021 10:40
Operator : SY/MD
Sample : VV1102WBL03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 72 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK240

Quant Time: Nov 07 01:02:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 02 23:31:16 2021
Response via : Initial Calibration

