

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110221\
 Data File : VV023135.D
 Acq On : 02 Nov 2021 10:03
 Operator : SY/MD
 Sample : VV1102WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK238

Quant Time: Nov 02 22:40:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 02 22:38:58 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	125794	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	124738	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	57945	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	51621	4.696	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.000%
7) Chloroethane-d5	1.564	69	39849	5.861	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	117.200%
11) 1,1-Dichloroethene-d2	2.104	63	68425	4.315	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.400%
20) 2-Butanone-d5	3.902	46	62616	35.511	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	71.020%
24) Chloroform-d	4.342	84	95291	5.333	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.600%
26) 1,2-Dichloroethane-d4	5.030	65	42790	5.080	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.600%
32) Benzene-d6	5.043	84	185130	5.084	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.600%
36) 1,2-Dichloropropane-d6	6.066	67	54931	4.900	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.000%
41) Toluene-d8	7.313	98	158111	4.833	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.600%
43) trans-1,3-Dichloroprop...	7.622	79	19505	4.966	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.400%
46) 2-Hexanone-d5	8.088	63	66429	45.679	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.360%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	35787	4.623	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	57570	5.568	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110221\
Data File : VV023135.D
Acq On : 02 Nov 2021 10:03
Operator : SY/MD
Sample : VV1102WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK238

Quant Time: Nov 02 22:40:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 02 22:38:58 2021
Response via : Initial Calibration

