

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV110320\
 Data File : VV019218.D
 Acq On : 02 Nov 2020 23:42
 Operator : SY/MD
 Sample : VV1103MBL02
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK308

Quant Time: Nov 03 06:17:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM110320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 03 05:28:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	242560	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	229758	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.27	152	115326	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	87690	46.84	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.68%
7) Chloroethane-d5	1.58	69	76323	52.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	104.06%
11) 1,1-Dichloroethene-d2	2.12	63	137827	39.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.54%
21) 2-Butanone-d5	3.91	46	112010	108.52	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.52%
24) Chloroform-d	4.38	84	171697	49.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.02%
26) 1,2-Dichloroethane-d4	5.06	65	122886	50.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.68%
32) Benzene-d6	5.08	84	325495	51.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.62%
36) 1,2-Dichloropropane-d6	6.09	67	100815	53.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.44%
41) Toluene-d8	7.34	98	295967	50.08	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.16%
43) trans-1,3-Dichloropropene-	7.64	79	53927	50.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.56%
47) 2-Hexanone-d5	8.11	63	74808	101.73	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.73%
56) 1,1,2,2-Tetrachloroethane-	10.23	84	124218	49.22	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.44%
66) 1,2-Dichlorobenzene-d4	11.65	152	124868	54.12	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.24%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV110320\
Data File : VV019218.D
Acq On : 02 Nov 2020 23:42
Operator : SY/MD
Sample : VV1103MBL02
Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK308

Quant Time: Nov 03 06:17:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM110320WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 03 05:28:09 2020
Response via : Initial Calibration

