

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110321\
 Data File : VV023188.D
 Acq On : 03 Nov 2021 15:46
 Operator : SY/MD
 Sample : VV1103WBL06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK276

Quant Time: Nov 08 21:07:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM102821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 30 00:25:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	251836	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	250286	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	125387	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	69838	40.239	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.480%
7) Chloroethane-d5	1.568	69	60589	44.758	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.520%
11) 1,1-Dichloroethene-d2	2.108	63	107932	32.830	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.660%
21) 2-Butanone-d5	3.879	46	131971	101.491	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.490%
24) Chloroform-d	4.349	84	176702	48.423	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.840%
26) 1,2-Dichloroethane-d4	5.034	65	113831	52.703	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.400%
32) Benzene-d6	5.050	84	337182	49.179	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.360%
36) 1,2-Dichloropropane-d6	6.069	67	106514	48.802	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.600%
41) Toluene-d8	7.317	98	301314	48.169	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.340%
43) trans-1,3-Dichloroprop...	7.622	79	51576	49.302	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.600%
47) 2-Hexanone-d5	8.088	63	95235	100.462	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.460%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	152810	47.902	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	131772	54.482	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110321\
Data File : VV023188.D
Acq On : 03 Nov 2021 15:46
Operator : SY/MD
Sample : VV1103WBL06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK276

Quant Time: Nov 08 21:07:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM102821WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 30 00:25:16 2021
Response via : Initial Calibration

