

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110321\
 Data File : VV023189.D
 Acq On : 03 Nov 2021 16:10
 Operator : SY/MD
 Sample : VV1103WBL07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK277

Quant Time: Nov 03 16:55:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM102821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 30 00:25:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	247679	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	244590	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	120910	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	69640	40.799	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.600%
7) Chloroethane-d5	1.568	69	57747	43.375	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.740%
11) 1,1-Dichloroethene-d2	2.108	63	106605	32.970	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.940%
21) 2-Butanone-d5	3.880	46	135176	105.701	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.700%
24) Chloroform-d	4.349	84	172219	47.987	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.980%
26) 1,2-Dichloroethane-d4	5.031	65	110900	52.208	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.420%
32) Benzene-d6	5.050	84	329805	49.223	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.440%
36) 1,2-Dichloropropane-d6	6.069	67	105681	49.548	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.100%
41) Toluene-d8	7.317	98	293425	48.001	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.000%
43) trans-1,3-Dichloroprop...	7.622	79	50022	48.930	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.860%
47) 2-Hexanone-d5	8.088	63	95519	103.108	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.110%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	149028	47.805	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.600%
66) 1,2-Dichlorobenzene-d4	11.625	152	128061	54.908	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.820%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110321\
Data File : VV023189.D
Acq On : 03 Nov 2021 16:10
Operator : SY/MD
Sample : VV1103WBL07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK277

Quant Time: Nov 03 16:55:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM102821WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 30 00:25:16 2021
Response via : Initial Calibration

