

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW110322\
 Data File : VW028849.D
 Acq On : 03 Nov 2022 15:31
 Operator : SY/MD
 Sample : N5421-01ME
 Misc : 2.92g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C1113ME

Quant Time: Nov 04 04:44:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 04 04:43:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.606	114	429287	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	394892	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	173591	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	142157	34.932	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.860%
7) Chloroethane-d5	1.555	69	138222	39.732	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.460%
11) 1,1-Dichloroethene-d2	2.082	63	219561	28.363	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	56.720%#
21) 2-Butanone-d5	3.889	46	263427	88.924	ug/L	0.03
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.920%
24) Chloroform-d	4.336	84	251182	40.924	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.840%
26) 1,2-Dichloroethane-d4	5.021	65	193779	47.737	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.480%
32) Benzene-d6	5.037	84	544612	44.178	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.360%
36) 1,2-Dichloropropane-d6	6.059	67	199238	47.076	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.160%
41) Toluene-d8	7.310	98	458164	42.276	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.560%
43) trans-1,3-Dichloroprop...	7.616	79	66049	40.633	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.260%
47) 2-Hexanone-d5	8.095	63	61819	32.322	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	32.320%#
56) 1,1,2,2-Tetrachloroeth...	10.217	84	251107	44.994	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.980%
66) 1,2-Dichlorobenzene-d4	11.622	152	159242	46.473	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.940%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110322\
Data File : VV028849.D
Acq On : 03 Nov 2022 15:31
Operator : SY/MD
Sample : N5421-01ME
Misc : 2.92g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C1113ME

Quant Time: Nov 04 04:44:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 04 04:43:26 2022
Response via : Initial Calibration

