




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110322\
 Data File : VV028850.D
 Acq On : 03 Nov 2022 16:11
 Operator : SY/MD
 Sample : N5421-02ME
 Misc : 4.79g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C1114ME

Manual IntegrationsAPPROVED

Quant Time: Nov 04 04:44:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 04 04:43:26 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 11/04/2022
 Supervised By :Mahesh Dadoda 11/04/2022



TIC: VV028850.D\data.ms

(47) 2-Hexanone-d5 (S)

8.091min (+ 0.013) 33.54 ug/L

response 71321

Ion	Exp%	Act%
63.00	100.00	100.00
46.00	191.30	170.73
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110322\
 Data File : VV028850.D
 Acq On : 03 Nov 2022 16:11
 Operator : SY/MD
 Sample : N5421-02ME
 Misc : 4.79g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C1114ME

Manual IntegrationsAPPROVED

Quant Time: Nov 04 04:44:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 04 04:43:26 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 11/04/2022
 Supervised By :Mahesh Dadoda 11/04/2022



TIC: VV028850.D\data.ms

(47) 2-Hexanone-d5 (S)

8.117min (+ 0.039) 83.20 ug/L m

response 176911

Ion	Exp%	Act%
63.00	100.00	100.00
46.00	191.30	68.83#
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110322\
 Data File : VV028850.D
 Acq On : 03 Nov 2022 16:11
 Operator : SY/MD
 Sample : N5421-02ME
 Misc : 4.79g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C1114ME

Manual IntegrationsAPPROVED

Quant Time: Nov 04 04:44:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Nov 04 04:43:26 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 11/04/2022
 Supervised By :Mahesh Dadoda 11/04/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.603	114	471679	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	439015	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	199752	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	140738	31.475	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	62.960%	
7) Chloroethane-d5	1.558	69	137850	36.064	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	72.120%	
11) 1,1-Dichloroethene-d2	2.082	63	215975	25.392	ug/L	-0.02
Spiked Amount 50.000	Range 60	- 125	Recovery	=	50.780%#	
21) 2-Butanone-d5	3.883	46	264869	81.375	ug/L	0.02
Spiked Amount 100.000	Range 40	- 130	Recovery	=	81.380%	
24) Chloroform-d	4.333	84	247726	36.733	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	73.460%	
26) 1,2-Dichloroethane-d4	5.018	65	194106	43.520	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	87.040%	
32) Benzene-d6	5.034	84	539404	39.358	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	78.720%	
36) 1,2-Dichloropropane-d6	6.056	67	195816	41.617	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	83.240%	
41) Toluene-d8	7.307	98	453880	37.672	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	75.340%#	
43) trans-1,3-Dichloroprop...	7.616	79	65372	36.175	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	72.340%	
47) 2-Hexanone-d5	8.117	63	176911m	83.201	ug/L	0.04
Spiked Amount 100.000	Range 45	- 130	Recovery	=	83.200%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	249854	40.270	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	80.540%	
66) 1,2-Dichlorobenzene-d4	11.619	152	163583	41.488	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	82.980%	
Target Compounds						
					Qvalue	
34) Trichloroethene	5.902	95	122127	33.789	ug/L	96
46) Tetrachloroethene	7.973	164	3050	1.218	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110322\
Data File : VV028850.D
Acq On : 03 Nov 2022 16:11
Operator : SY/MD
Sample : N5421-02ME
Misc : 4.79g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C1114ME

Manual IntegrationsAPPROVED

Quant Time: Nov 04 04:44:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103122WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Nov 04 04:43:26 2022
Response via : Initial Calibration

Reviewed By :Krupa Patel 11/04/2022
Supervised By :Mahesh Dadoda 11/04/2022

