

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110323\
 Data File : VV032833.D
 Acq On : 02 Nov 2023 09:43
 Operator : SY/MD
 Sample : VV1103WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK217

Quant Time: Nov 03 03:25:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101323.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Oct 16 02:50:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	4945	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.800	117	4768	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2474	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	4193	0.517	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	104.000%
4) 1,2-Dichloroethane-d4	4.946	65	2710	0.528	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.000%
7) 1,2-Dichloropropane-d6	5.995	67	2978	0.455	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	90.000%
8) Toluene-d8	7.253	98	7288	0.469	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	94.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2112	0.469	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	94.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110323\
Data File : VV032833.D
Acq On : 02 Nov 2023 09:43
Operator : SY/MD
Sample : VV1103WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK217

Quant Time: Nov 03 03:25:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101323.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Oct 16 02:50:34 2023
Response via : Initial Calibration

