

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110323\
 Data File : VV032834.D
 Acq On : 02 Nov 2023 10:12
 Operator : SY/MD
 Sample : 05168-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 XA013

Quant Time: Nov 03 03:26:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101323.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Oct 16 02:50:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	5112	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.800	117	4995	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2670	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	4148	0.495	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	100.000%
4) 1,2-Dichloroethane-d4	4.946	65	2849	0.537	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	108.000%
7) 1,2-Dichloropropane-d6	5.995	67	3188	0.465	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	92.000%
8) Toluene-d8	7.253	98	7664	0.470	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	94.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2282	0.484	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	96.000%
Target Compounds						
					Qvalue	
9) 1,2-Dibromoethane	8.290	107	2200	0.634	ug/L	100
12) 1,2,3-Trichloropropane	10.197	75	2844	0.924	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.356	75	633	1.185	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110323\
Data File : VV032834.D
Acq On : 02 Nov 2023 10:12
Operator : SY/MD
Sample : 05168-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
XA013

Quant Time: Nov 03 03:26:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101323.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Oct 16 02:50:34 2023
Response via : Initial Calibration

