

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110323\
 Data File : VV032835.D
 Acq On : 02 Nov 2023 10:50
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5307

Quant Time: Nov 03 03:29:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101323.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Oct 16 02:50:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	6494	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.800	117	6340	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	3273	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	5799	0.545	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 108.000%	
4) 1,2-Dichloroethane-d4	4.946	65	3863	0.573	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 114.000%	
7) 1,2-Dichloropropane-d6	5.996	67	4392	0.504	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 100.000%	
8) Toluene-d8	7.238	98	10709	0.518	ug/L	-0.01
Spiked Amount	0.500	Range	70 - 130	Recovery	= 104.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3100	0.518	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 104.000%	
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.280	62	3922	0.508	ug/L	99
6) Trichloroethene	5.829	95	3249	0.478	ug/L	92
9) 1,2-Dibromoethane	8.290	107	2074	0.471	ug/L	98
12) 1,2,3-Trichloropropane	10.197	75	1878	0.498	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.357	75	319	0.487	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110323\
Data File : VV032835.D
Acq On : 02 Nov 2023 10:50
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5307

Quant Time: Nov 03 03:29:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101323.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Oct 16 02:50:34 2023
Response via : Initial Calibration

