

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV110419\
 Data File : VV013478.D
 Acq On : 04 Nov 2019 16:40
 Operator : SY/MD
 Sample : K5635-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Nov 05 07:26:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR103119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 05 07:17:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	337897	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	354995	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	179024	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	129628	5.21	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.20%
7) Chloroethane-d5	1.58	69	112357	5.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.40%
11) 1,1-Dichloroethene-d2	2.13	63	179244	4.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.60%
20) 2-Butanone-d5	3.96	46	382275	49.55	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.10%
24) Chloroform-d	4.40	84	301032	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.80%
26) 1,2-Dichloroethane-d4	5.08	65	162357	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.60%
32) Benzene-d6	5.09	84	613816	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.60%
36) 1,2-Dichloropropane-d6	6.11	67	191790	5.21	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.20%
41) Toluene-d8	7.35	98	512064	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
43) trans-1,3-Dichloropropene-	7.66	79	66068	4.47	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.40%
46) 2-Hexanone-d5	8.13	63	211425	38.93	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.86%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	120660	4.27	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	198799	5.46	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.20%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV110419\
Data File : VV013478.D
Acq On : 04 Nov 2019 16:40
Operator : SY/MD
Sample : K5635-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Nov 05 07:26:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR103119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 05 07:17:32 2019
Response via : Initial Calibration

