

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV110420\
 Data File : VV019229.D
 Acq On : 03 Nov 2020 12:02
 Operator : SY/MD
 Sample : VV1104MBL01
 Misc : 5.0uL/5.0mL/100uL/5.1mL/MSVOA_V/MEOH
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK310

Quant Time: Nov 04 07:17:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM110320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 04 07:03:57 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	261087	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	252033	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.27	152	129045	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	90523	44.93	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.86%
7) Chloroethane-d5	1.58	69	76460	48.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.86%
11) 1,1-Dichloroethene-d2	2.12	63	144691	38.79	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.58%
21) 2-Butanone-d5	3.91	46	113554	102.21	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.21%
24) Chloroform-d	4.38	84	174184	46.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.38%
26) 1,2-Dichloroethane-d4	5.05	65	120495	45.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.72%
32) Benzene-d6	5.07	84	328614	47.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.44%
36) 1,2-Dichloropropane-d6	6.09	67	100780	48.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.00%
41) Toluene-d8	7.34	98	302868	46.72	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.44%
43) trans-1,3-Dichloropropene-	7.64	79	54137	46.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.04%
47) 2-Hexanone-d5	8.11	63	74362	92.19	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.19%
56) 1,1,2,2-Tetrachloroethane-	10.23	84	122389	44.21	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.42%
66) 1,2-Dichlorobenzene-d4	11.65	152	127660	49.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.90%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV110420\
Data File : VV019229.D
Acq On : 03 Nov 2020 12:02
Operator : SY/MD
Sample : VV1104MBL01
Misc : 5.0µL/5.0mL/100µL/5.1mL/MSVOA_V/MEOH
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK310

Quant Time: Nov 04 07:17:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM110320WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 04 07:03:57 2020
Response via : Initial Calibration

