

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110421\
 Data File : VV023203.D
 Acq On : 04 Nov 2021 15:08
 Operator : SY/MD
 Sample : VV1104WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK241

Quant Time: Nov 09 01:06:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 08 13:19:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	139505	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	137184	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	65128	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	44024	5.037	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 100.800%		
7) Chloroethane-d5	1.568	69	37585	5.277	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 105.600%		
11) 1,1-Dichloroethene-d2	2.111	63	61682	3.770	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 75.400%		
20) 2-Butanone-d5	3.889	46	91306	60.642	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	= 121.280%		
24) Chloroform-d	4.352	84	91512	4.913	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 98.200%		
26) 1,2-Dichloroethane-d4	5.034	65	43054	5.141	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 102.800%		
32) Benzene-d6	5.053	84	176639	5.018	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 100.400%		
36) 1,2-Dichloropropane-d6	6.072	67	53553	5.168	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 103.400%		
41) Toluene-d8	7.317	98	154340	4.679	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 93.600%		
43) trans-1,3-Dichloroprop...	7.625	79	19095	4.860	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 97.200%		
46) 2-Hexanone-d5	8.088	63	72706	50.296	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 100.600%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	35936	4.823	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 96.400%		
66) 1,2-Dichlorobenzene-d4	11.625	152	59811	5.515	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 110.400%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110421\
Data File : VV023203.D
Acq On : 04 Nov 2021 15:08
Operator : SY/MD
Sample : VV1104WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK241

Quant Time: Nov 09 01:06:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 08 13:19:52 2021
Response via : Initial Calibration

