

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW110423\
 Data File : VW032844.D
 Acq On : 03 Nov 2023 11:46
 Operator : SY/MD
 Sample : 05226-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCG85

Quant Time: Nov 04 00:59:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Nov 04 00:57:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	345392	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	381834	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	169332	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	120280	36.613	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	73.220%	
7) Chloroethane-d5	1.532	69	110871	40.353	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	80.700%	
11) 1,1-Dichloroethene-d2	2.063	65	53311	39.289	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	78.580%	
21) 2-Butanone-d5	3.786	46	126727	88.935	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	88.940%	
24) Chloroform-d	4.249	84	231633	44.174	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	88.340%	
26) 1,2-Dichloroethane-d4	4.944	65	150826	46.280	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	92.560%	
32) Benzene-d6	4.963	84	448710	45.705	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.400%	
36) 1,2-Dichloropropane-d6	5.989	67	150985	50.621	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	101.240%	
41) Toluene-d8	7.246	98	422876	43.745	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	87.480%	
43) trans-1,3-Dichloroprop...	7.555	79	67560	43.492	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	86.980%	
47) 2-Hexanone-d5	8.027	63	94068	86.390	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	86.390%	
56) 1,1,2,2-Tetrachloroeth...	10.153	84	202704	44.786	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	89.580%	
66) 1,2-Dichlorobenzene-d4	11.561	152	170041	52.110	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	104.220%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110423\
Data File : VV032844.D
Acq On : 03 Nov 2023 11:46
Operator : SY/MD
Sample : 05226-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCG85

Quant Time: Nov 04 00:59:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Nov 04 00:57:36 2023
Response via : Initial Calibration

