

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV110518\
 Data File : VV008518.D
 Acq On : 05 Nov 2018 14:07
 Operator : SY/MD
 Sample : VSTD1.043
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD1.043

Quant Time: Nov 06 10:35:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR110518SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 06 10:32:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	7617	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	6658	0.50	ug/L	0.00
10) 1,4-Dichlorobenzene-d4	11.29	152	2836	0.50	ug/L	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) Vinyl Chloride-d3	1.32	65	6381	0.94	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.07	65	5576	1.02	ug/L	-0.02
7) Toluene-d8	7.36	98	17401	1.04	ug/L	0.01
9) 1,1,2,2-Tetrachloroethane-	10.25	84	4499	1.01	ug/L	0.00
11) 1-2-dichlorobenzene-d4	11.67	152	5834	1.02	ug/L	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Vinyl chloride	1.32	62	6110	0.97	ug/L #	31
6) Trichloroethene	5.96	95	6311	0.97	ug/L	99
8) 1,2-Dibromoethane	8.40	107	3624	0.97	ug/L #	99
12) 1,2,3-Trichloropropane	10.32	75	3475	1.00	ug/L	98
13) 1,2-Dibromo-3-chloropropan	12.46	75	674	0.94	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV110518\
Data File : VV008518.D
Acq On : 05 Nov 2018 14:07
Operator : SY/MD
Sample : VSTD1.043
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD1.043

Quant Time: Nov 06 10:35:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR110518SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 06 10:32:16 2018
Response via : Initial Calibration

