

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV110518\
 Data File : VV008519.D
 Acq On : 05 Nov 2018 14:36
 Operator : SY/MD
 Sample : VSTD2.044
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD2.044

Quant Time: Nov 06 10:36:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR110518SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 06 10:32:16 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	6843	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	5961	0.50	ug/L	0.00
10) 1,4-Dichlorobenzene-d4	11.29	152	2599	0.50	ug/L	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) Vinyl Chloride-d3	1.32	65	11270	1.84	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.09	65	10227	2.07	ug/L	0.00
7) Toluene-d8	7.35	98	30943	2.06	ug/L	0.00
9) 1,1,2,2-Tetrachloroethane-	10.25	84	8506	2.13	ug/L	0.00
11) 1-2-dichlorobenzene-d4	11.67	152	10669	2.05	ug/L	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Vinyl chloride	1.32	62	11244	1.98	ug/L #	25
6) Trichloroethene	5.96	95	11413	1.95	ug/L	99
8) 1,2-Dibromoethane	8.40	107	6907	2.06	ug/L #	97
12) 1,2,3-Trichloropropane	10.32	75	6662	2.10	ug/L	99
13) 1,2-Dibromo-3-chloropropan	12.46	75	1340	2.03	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV110518\
Data File : VV008519.D
Acq On : 05 Nov 2018 14:36
Operator : SY/MD
Sample : VSTD2.044
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD2.044

Quant Time: Nov 06 10:36:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR110518SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 06 10:32:16 2018
Response via : Initial Calibration

