

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV110518\
 Data File : VV008529.D
 Acq On : 05 Nov 2018 20:15
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0 mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.567

Quant Time: Nov 06 11:02:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR110518SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Nov 06 10:52:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	6500	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	5603	0.50	ug/L	0.00
10) 1,4-Dichlorobenzene-d4	11.29	152	2310	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.32	65	2758	0.53	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	106.00%
4) 1,2-Dichloroethane-d4	5.09	65	2460	0.52	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	104.00%
7) Toluene-d8	7.36	98	7179	0.50	ug/L	0.01
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.00%
9) 1,1,2,2-Tetrachloroethane-	10.25	84	1967	0.51	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.00%
11) 1-2-dichlorobenzene-d4	11.67	152	2372	0.49	ug/L	0.00
Spiked Amount	0.500		Recovery	=	98.00%	

Target Compounds

						Qvalue
3) Vinyl chloride	1.32	62	2823	0.52	ug/L	# 30
6) Trichloroethene	5.96	95	2791	0.48	ug/L	99
8) 1,2-Dibromoethane	8.40	107	1648	0.54	ug/L	100
12) 1,2,3-Trichloropropane	10.32	75	1553	0.57	ug/L	100
13) 1,2-Dibromo-3-chloropropan	12.48	75	286	0.47	ug/L	# 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV110518\
Data File : VV008529.D
Acq On : 05 Nov 2018 20:15
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25.0 mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.567

Quant Time: Nov 06 11:02:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR110518SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Nov 06 10:52:01 2018
Response via : Initial Calibration

