

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV110520\
 Data File : VV019254.D
 Acq On : 04 Nov 2020 17:12
 Operator : SY/MD
 Sample : L4615-10ME 4X
 Misc : 5.64g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG482ME

Quant Time: Nov 05 04:14:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Nov 05 02:20:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	256034	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	247961	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	127812	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	62042	30.49	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	60.98%
7) Chloroethane-d5	1.57	69	56597	35.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	71.62%
11) 1,1-Dichloroethene-d2	2.11	63	105508	27.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	54.52%#
21) 2-Butanone-d5	3.91	46	104111	83.54	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.54%
24) Chloroform-d	4.37	84	143554	37.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.86%
26) 1,2-Dichloroethane-d4	5.05	65	106549	40.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.80%
32) Benzene-d6	5.07	84	273958	38.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.90%
36) 1,2-Dichloropropane-d6	6.09	67	85951	40.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	81.96%
41) Toluene-d8	7.33	98	252840	37.42	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	74.84%#
43) trans-1,3-Dichloropropene-	7.64	79	45180	37.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.46%
47) 2-Hexanone-d5	8.11	63	72844	80.20	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.20%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	109438	38.70	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	77.40%
64) 1,2-Dichlorobenzene-d4	11.65	152	112964	42.07	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.14%


Target Compounds					Ovalue
46) Tetrachloroethene	8.00	164	40763	24.374 ug/L	98




(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV110520\
Data File : VV019254.D
Acq On : 04 Nov 2020 17:12
Operator : SY/MD
Sample : L4615-10ME 4X
Misc : 5.64g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG482ME

Quant Time: Nov 05 04:14:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Nov 05 02:20:23 2020
Response via : Initial Calibration





#46

Tetrachloroethene

Concen: 24.374 ug/L

RT: 8.00 min Scan# 2148

Delta R.T. 0.00 min

Lab File: VV019254.D

Acq: 04 Nov 2020 17:12

Instrument :

MSVOA_V

ClientSampleId :

BG482ME

Tot Ion:164 Resp: 40763

Ion Ratio Lower Upper

164 100

129 91.3 63.2 117.4

131 89.0 60.3 112.1

166 127.2 88.3 163.9

Abundance Ion 164.00 (163.70 to 164.70): V

Ion 129.00 (128.70 to 129.70): V

Ion 131.00 (130.70 to 131.70): V

Ion 166.00 (165.70 to 166.70): V

