

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW110521\
 Data File : VW023244.D
 Acq On : 05 Nov 2021 19:54
 Operator : SY/MD
 Sample : M4535-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H4626

Quant Time: Nov 09 03:40:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 09 02:04:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	131162	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	132097	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	61575	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	40276	4.902	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	98.000%	
7) Chloroethane-d5	1.568	69	34659	5.175	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	103.600%	
11) 1,1-Dichloroethene-d2	2.111	63	57532	3.740	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	74.800%	
20) 2-Butanone-d5	3.902	46	92918	65.638	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	131.280%#	
24) Chloroform-d	4.352	84	88571	5.058	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.200%	
26) 1,2-Dichloroethane-d4	5.037	65	41517	5.272	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	105.400%	
32) Benzene-d6	5.053	84	170487	5.030	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.600%	
36) 1,2-Dichloropropane-d6	6.072	67	52450	5.257	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	105.200%	
41) Toluene-d8	7.317	98	145636	4.585	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	91.800%	
43) trans-1,3-Dichloroprop...	7.628	79	16610	4.390	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	87.800%	
46) 2-Hexanone-d5	8.091	63	69025	49.589	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	99.180%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	34853	4.858	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	97.200%	
66) 1,2-Dichlorobenzene-d4	11.625	152	54513	5.317	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.400%	
Target Compounds						
					Qvalue	
10) 1,1,2-Trichloro-1,2,2-...	2.111	101	1142	0.139	ug/L #	58
12) 1,1-Dichloroethene	2.117	96	1600	0.205	ug/L #	1
17) Methyl tert-butyl Ether	2.773	73	2591	0.150	ug/L	96
22) cis-1,2-Dichloroethene	3.918	96	1370	0.148	ug/L #	86
34) Trichloroethene	5.921	95	13907	1.416	ug/L	97
47) Tetrachloroethene	7.976	164	112696	13.244	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110521\
Data File : VV023244.D
Acq On : 05 Nov 2021 19:54
Operator : SY/MD
Sample : M4535-10
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H4626

Quant Time: Nov 09 03:40:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 09 02:04:24 2021
Response via : Initial Calibration

