

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110521\
 Data File : VV023250.D
 Acq On : 05 Nov 2021 22:38
 Operator : SY/MD
 Sample : M4535-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H4597

Quant Time: Nov 09 04:09:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 09 03:48:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	126874	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	126711	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	58833	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	36564	4.600	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.000%
7) Chloroethane-d5	1.568	69	32139	4.961	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.200%
11) 1,1-Dichloroethene-d2	2.111	63	53395	3.589	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.800%
20) 2-Butanone-d5	3.895	46	82492	60.243	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.480%
24) Chloroform-d	4.349	84	81944	4.838	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
26) 1,2-Dichloroethane-d4	5.037	65	38760	5.089	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.800%
32) Benzene-d6	5.053	84	153412	4.719	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.400%
36) 1,2-Dichloropropane-d6	6.072	67	47709	4.985	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.800%
41) Toluene-d8	7.317	98	133015	4.366	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
43) trans-1,3-Dichloroprop...	7.625	79	16831	4.638	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.800%
46) 2-Hexanone-d5	8.088	63	58612	43.898	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.800%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	32531	4.727	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.600%
66) 1,2-Dichlorobenzene-d4	11.625	152	50916	5.197	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.000%
Target Compounds					Qvalue	
13) Acetone	2.188	43	2368	2.830	ug/L	73
22) cis-1,2-Dichloroethene	3.915	96	3425	0.383	ug/L #	90
34) Trichloroethene	5.924	95	4160	0.442	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110521\
Data File : VV023250.D
Acq On : 05 Nov 2021 22:38
Operator : SY/MD
Sample : M4535-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H4597

Quant Time: Nov 09 04:09:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 09 03:48:20 2021
Response via : Initial Calibration

