

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW110521\
 Data File : VW023252.D
 Acq On : 05 Nov 2021 23:25
 Operator : SY/MD
 Sample : M4535-17
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H4600

Quant Time: Nov 09 04:14:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 09 03:48:20 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	130281	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	129723	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	59323	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	37706	4.620	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	92.400%	
7) Chloroethane-d5	1.568	69	33871	5.092	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	101.800%	
11) 1,1-Dichloroethene-d2	2.111	63	54687	3.579	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	71.600%	
20) 2-Butanone-d5	3.892	46	88035	62.609	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	125.220%	
24) Chloroform-d	4.352	84	85632	4.923	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.400%	
26) 1,2-Dichloroethane-d4	5.037	65	41115	5.257	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	105.200%	
32) Benzene-d6	5.053	84	160196	4.813	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.200%	
36) 1,2-Dichloropropane-d6	6.072	67	50116	5.115	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	102.200%	
41) Toluene-d8	7.317	98	136250	4.368	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.400%	
43) trans-1,3-Dichloroprop...	7.629	79	16920	4.554	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	91.000%	
46) 2-Hexanone-d5	8.088	63	67445	49.340	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	98.680%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	35180	4.993	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	99.800%	
66) 1,2-Dichlorobenzene-d4	11.625	152	53999	5.467	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	109.400%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.134	85	13457	1.059	ug/L	96
9) Trichlorofluoromethane	1.754	101	1861	0.115	ug/L	89
13) Acetone	2.182	43	1557	1.812	ug/L	72
25) Chloroform	4.378	83	7820	0.455	ug/L	99
34) Trichloroethene	5.928	95	3423	0.355	ug/L	88
38) Bromodichloromethane	6.519	83	1064	0.094	ug/L #	79
47) Tetrachloroethene	7.976	164	43387	5.192	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110521\
Data File : VV023252.D
Acq On : 05 Nov 2021 23:25
Operator : SY/MD
Sample : M4535-17
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H4600

Quant Time: Nov 09 04:14:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 09 03:48:20 2021
Response via : Initial Calibration

