

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110525\
 Data File : VV039337.D
 Acq On : 05 Nov 2025 15:50
 Operator : SY/MD
 Sample : Q3552-06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 TB

Quant Time: Nov 06 00:16:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:01:26 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.596	168	856878	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.529	114	1636404	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.773	117	1498704	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	661803	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.934	65	703372	57.768	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	115.540%
35) Dibromofluoromethane	4.497	113	623292	48.341	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	96.680%
50) Toluene-d8	7.233	98	2056521	46.884	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	93.760%
62) 4-Bromofluorobenzene	9.998	95	679949	43.526	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	87.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110525\
Data File : VV039337.D
Acq On : 05 Nov 2025 15:50
Operator : SY/MD
Sample : Q3552-06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
TB

Quant Time: Nov 06 00:16:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:01:26 2025
Response via : Initial Calibration

