

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110525\
 Data File : VV039343.D
 Acq On : 05 Nov 2025 18:05
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK

Quant Time: Nov 06 00:19:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 08 05:01:26 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	587692	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	1063324	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.773	117	1010398	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.172	152	449521	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.941	65	529468	63.403	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	126.800%#
35) Dibromofluoromethane	4.500	113	477139	56.950	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	113.900%
50) Toluene-d8	7.236	98	1508587	52.928	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	105.860%
62) 4-Bromofluorobenzene	9.998	95	516782	50.910	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	101.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110525\
Data File : VV039343.D
Acq On : 05 Nov 2025 18:05
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK

Quant Time: Nov 06 00:19:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V100725W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 08 05:01:26 2025
Response via : Initial Calibration

