


Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV110619\
 Data File : VV013489.D
 Acq On : 06 Nov 2019 09:44
 Operator : SY/MD
 Sample : BFB72
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB72

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR110419WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Tue Nov 05 11:12:11 2019



AutoFind: Scans 2806, 2807, 2808; Background Corrected with Scan 2790

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	11258	PASS
75	95	30	80	49.9	29032	PASS
95	95	100	100	100.0	58170	PASS
96	95	5	9	7.0	4046	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	95.9	55802	PASS
175	174	5	9	6.8	3799	PASS
176	174	95	101	96.6	53893	PASS
177	176	5	9	6.6	3567	PASS

m/z	Abundance
36.00	512.0
37.00	2905.0
38.00	2511.0
39.00	1175.0
39.90	1028.0
41.20	142.0
41.90	148.0
42.60	148.0
43.10	133.0
44.00	1842.0
45.20	663.0
45.90	176.0
47.00	971.0
48.10	463.0
49.00	2230.0
50.10	11994.0
51.10	3448.0
51.90	167.0
53.20	252.0
53.80	146.0
54.20	201.0
55.00	293.0
56.00	725.0
57.00	1577.0
58.10	186.0
59.90	412.0
61.00	3232.0
62.10	2990.0
63.00	2563.0
63.90	411.0
65.00	1028.0
65.90	314.0
67.20	205.0
68.00	5830.0
69.00	6498.0
70.10	524.0
72.10	531.0
73.00	2989.0
74.10	9642.0
75.00	30056.0
76.00	2398.0
76.90	513.0
78.10	386.0
78.90	1583.0
79.80	688.0
80.90	2094.0
81.90	500.0
84.80	113.0
86.00	103.0
86.90	2569.0
87.90	2388.0
91.10	406.0
92.00	1901.0
93.10	2783.0
94.00	7281.0
95.00	60256.0
96.00	4156.0
97.10	229.0
103.90	364.0
105.30	139.0
105.90	260.0
107.10	221.0
107.90	103.0
109.60	151.0
109.90	173.0
110.50	116.0
112.90	114.0
114.80	108.0
115.80	310.0
116.90	515.0
117.80	341.0
119.00	377.0
123.70	118.0
125.80	141.0
126.60	103.0
127.70	247.0
128.20	156.0
128.90	250.0
129.90	250.0
131.10	145.0
134.10	113.0
135.00	165.0
135.80	148.0
136.80	144.0
137.60	118.0
140.90	562.0
141.70	138.0
142.90	999.0
143.80	175.0
145.00	486.0
147.80	234.0
149.10	101.0

Instrument :
MSVOA_V
ClientSampleId :
BFB72

149.90	179.0
152.80	263.0
154.70	259.0
155.90	131.0
157.10	272.0
158.80	190.0
160.70	149.0
162.70	123.0
168.30	105.0
170.90	329.0
172.00	953.0
173.90	57656.0
175.00	3794.0
176.00	55936.0
176.90	3670.0
178.10	162.0
189.50	172.0
205.10	170.0
206.90	101.0
213.10	112.0
221.20	107.0
239.00	107.0
258.90	108.0
293.80	103.0

Instrument :
MSVOA_V
ClientSampleId :
BFB72

m/z	Abundance
36.00	578.0
37.00	2690.0
38.00	2388.0
39.00	1149.0
39.90	1018.0
41.20	185.0
41.90	178.0
42.60	131.0
43.10	206.0
44.00	1657.0
45.20	762.0
46.00	196.0
47.00	875.0
48.10	416.0
49.00	2092.0
50.10	10960.0
51.10	3496.0
52.00	230.0
53.30	276.0
53.80	137.0
54.20	219.0
54.90	295.0
56.00	692.0
57.10	1399.0
58.00	179.0
60.00	437.0
61.00	3059.0
62.10	3017.0
63.00	2428.0
64.00	360.0
65.00	818.0
65.90	244.0
68.00	5224.0
69.00	6253.0
70.10	514.0
71.10	131.0
72.10	600.0
73.00	2899.0
74.10	9010.0
75.00	28176.0
76.00	2432.0
77.00	444.0
78.10	345.0
78.90	1562.0
79.90	769.0
80.90	2098.0
81.90	394.0
86.90	2532.0
87.90	2412.0
91.10	346.0
92.00	1838.0
93.10	2597.0
94.00	6917.0
95.00	55240.0
96.00	4143.0
97.10	280.0
103.90	360.0
105.30	184.0
105.80	212.0
107.10	204.0
107.80	102.0
109.60	153.0
110.60	151.0
112.90	127.0
115.80	253.0
116.10	237.0
116.90	467.0
117.80	382.0
118.30	181.0
119.00	393.0
123.70	143.0
125.70	142.0
126.60	103.0
127.70	191.0
128.10	133.0
128.90	305.0
129.80	274.0
131.00	160.0
132.10	123.0
134.00	107.0
134.90	175.0
135.80	138.0
136.80	139.0
140.80	551.0
141.70	137.0
142.90	902.0
143.80	128.0
145.10	347.0
145.80	109.0
147.90	262.0
149.10	158.0
149.80	146.0

Instrument :
MSVOA_V
ClientSampleId :
BFB72

152.80	220.0
154.80	235.0
156.90	251.0
158.80	183.0
160.70	117.0
162.70	154.0
168.30	119.0
170.90	310.0
172.10	862.0
173.90	52000.0
175.00	3605.0
176.00	50984.0
176.90	3205.0
177.80	182.0
178.20	170.0
188.00	129.0
189.50	118.0
205.10	154.0
249.80	137.0
274.10	107.0

Instrument :
MSVOA_V
ClientSampleId :
BFB72

m/z	Abundance
35.60	169.0
36.30	173.0
37.20	192.0
37.90	127.0
38.40	213.0
39.00	146.0
39.90	1052.0
41.10	191.0
41.60	170.0
42.30	105.0
44.00	1211.0
45.80	165.0
46.10	173.0
46.80	126.0
47.60	125.0
48.30	123.0
49.60	179.0
50.20	202.0
51.20	107.0
52.00	126.0
52.70	104.0
53.50	161.0
54.10	113.0
55.30	180.0
56.00	133.0
57.60	193.0
58.50	184.0
62.30	135.0
63.70	138.0
64.10	140.0
67.40	111.0
68.40	111.0
70.10	107.0
71.70	129.0
72.70	135.0
76.70	138.0
78.00	144.0
79.30	128.0
81.70	116.0
83.90	156.0
85.50	170.0
89.80	102.0
91.00	106.0
92.60	113.0
99.20	128.0
100.20	155.0
102.80	105.0
104.70	101.0
107.00	155.0
108.80	113.0
117.10	151.0
121.50	130.0
121.90	113.0
129.70	142.0
133.50	115.0
136.80	106.0
149.80	151.0
154.40	107.0
157.30	121.0
171.80	117.0
179.40	107.0
193.60	106.0
226.10	116.0
228.00	112.0
237.90	108.0
240.20	102.0
269.30	145.0

Instrument :
MSVOA_V
ClientSampleId :
BFB72

m/z	Abundance
36.10	443.0
37.00	2927.0
38.00	2501.0
39.00	1105.0
39.90	1195.0
41.30	146.0
42.10	154.0
42.50	167.0
44.00	1716.0
45.10	542.0
45.90	151.0
47.00	919.0
48.10	462.0
49.10	2134.0
50.10	11427.0
51.00	3234.0
51.90	135.0
53.20	242.0
53.80	176.0
55.00	293.0
56.00	700.0
57.00	1495.0
58.10	133.0
59.90	396.0
61.00	2932.0
62.00	2748.0
63.00	2381.0
64.00	502.0
65.00	1039.0
65.90	285.0
67.10	194.0
68.00	5838.0
69.00	6195.0
70.10	472.0
72.00	462.0
73.10	2801.0
74.10	9315.0
75.00	28864.0
76.00	2104.0
76.90	493.0
78.10	401.0
78.90	1553.0
79.80	581.0
80.90	1879.0
81.90	560.0
83.00	158.0
84.80	112.0
86.00	148.0
87.00	2509.0
88.00	2294.0
89.10	118.0
91.10	365.0
92.00	1745.0
93.00	2579.0
94.00	6974.0
95.00	59016.0
96.00	3841.0
97.10	167.0
104.00	372.0
104.70	117.0
105.90	326.0
107.20	174.0
109.60	124.0
110.00	189.0
114.90	141.0
115.90	292.0
116.80	521.0
117.90	267.0
119.00	360.0
124.90	112.0
125.90	117.0
127.70	247.0
128.20	179.0
128.90	138.0
130.00	255.0
131.10	136.0
134.50	132.0
135.00	147.0
135.80	106.0
136.80	148.0
137.60	104.0
140.90	592.0
141.70	110.0
142.90	966.0
143.80	188.0
145.00	533.0
146.10	123.0
147.70	189.0
149.90	190.0
153.00	213.0
154.60	260.0
155.90	137.0

Instrument :
MSVOA_V
ClientSampleId :
BFB72

157.20	274.0
158.80	151.0
160.70	147.0
170.90	286.0
172.00	1063.0
173.90	57752.0
175.00	3998.0
176.00	54760.0
177.00	3826.0
178.10	157.0
187.40	109.0
189.50	170.0
195.40	109.0
205.10	153.0
206.90	122.0
213.00	102.0
239.00	119.0
258.90	129.0

Instrument :
MSVOA_V
ClientSampleId :
BFB72