

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV110620\
 Data File : VV019258.D
 Acq On : 04 Nov 2020 21:34
 Operator : SY/MD
 Sample : VV1106WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK314

Quant Time: Nov 05 05:10:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM110320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Nov 05 04:58:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	241405	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	236470	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.27	152	115954	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	81616	43.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.62%
7) Chloroethane-d5	1.58	69	69285	47.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.92%
11) 1,1-Dichloroethene-d2	2.12	63	126593	36.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.40%
21) 2-Butanone-d5	3.91	46	110958	108.01	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.01%
24) Chloroform-d	4.38	84	160392	46.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.00%
26) 1,2-Dichloroethane-d4	5.06	65	112580	46.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.68%
32) Benzene-d6	5.07	84	298147	45.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.32%
36) 1,2-Dichloropropane-d6	6.09	67	93135	47.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.54%
41) Toluene-d8	7.34	98	266883	43.88	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.76%
43) trans-1,3-Dichloropropene-	7.64	79	47652	43.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.34%
47) 2-Hexanone-d5	8.11	63	69501	91.83	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.83%
56) 1,1,2,2-Tetrachloroethane-	10.23	84	117199	45.12	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.24%
66) 1,2-Dichlorobenzene-d4	11.65	152	113250	48.82	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.64%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV110620\
Data File : VV019258.D
Acq On : 04 Nov 2020 21:34
Operator : SY/MD
Sample : VV1106WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK314

Quant Time: Nov 05 05:10:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM110320WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Nov 05 04:58:48 2020
Response via : Initial Calibration

