

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV110620\
 Data File : VV019261.D
 Acq On : 04 Nov 2020 23:09
 Operator : SY/MD
 Sample : VV1106MBL02
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK317

Quant Time: Nov 05 05:18:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM110320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Nov 05 04:58:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	234675	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	227126	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.27	152	110263	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	78127	43.14	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.28%
7) Chloroethane-d5	1.58	69	68846	48.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.02%
11) 1,1-Dichloroethene-d2	2.12	63	119620	35.67	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.34%
21) 2-Butanone-d5	3.91	46	112436	112.59	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	112.59%
24) Chloroform-d	4.38	84	156059	46.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.08%
26) 1,2-Dichloroethane-d4	5.06	65	114065	48.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.60%
32) Benzene-d6	5.07	84	294581	46.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.94%
36) 1,2-Dichloropropane-d6	6.09	67	93750	50.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.14%
41) Toluene-d8	7.34	98	259072	44.34	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.68%
43) trans-1,3-Dichloropropene-	7.64	79	47457	44.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.52%
47) 2-Hexanone-d5	8.11	63	68314	93.98	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	93.98%
56) 1,1,2,2-Tetrachloroethane-	10.23	84	115401	46.25	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.50%
66) 1,2-Dichlorobenzene-d4	11.65	152	111670	50.62	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.24%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV110620\
Data File : VV019261.D
Acq On : 04 Nov 2020 23:09
Operator : SY/MD
Sample : VV1106MBL02
Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK317

Quant Time: Nov 05 05:18:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVLM110320WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Nov 05 04:58:48 2020
Response via : Initial Calibration

