

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW110623\
 Data File : VW032854.D
 Acq On : 06 Nov 2023 15:35
 Operator : SY/MD
 Sample : 05259-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YCGA9

Quant Time: Nov 07 01:36:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Nov 07 01:35:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	341181	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	380274	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	170755	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	107757	33.206	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	66.420%	
7) Chloroethane-d5	1.532	69	103326	38.071	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	76.140%	
11) 1,1-Dichloroethene-d2	2.059	65	50062	37.350	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	74.700%	
21) 2-Butanone-d5	3.789	46	128642	91.393	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	91.390%	
24) Chloroform-d	4.252	84	219616	42.399	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	84.800%	
26) 1,2-Dichloroethane-d4	4.944	65	140984	43.794	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.580%	
32) Benzene-d6	4.963	84	407248	41.652	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	83.300%	
36) 1,2-Dichloropropane-d6	5.989	67	142508	47.975	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	95.960%	
41) Toluene-d8	7.246	98	385040	39.994	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	79.980%#	
43) trans-1,3-Dichloroprop...	7.554	79	63901	41.306	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	82.620%	
47) 2-Hexanone-d5	8.027	63	93247	85.988	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	85.990%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	202863	45.005	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	90.000%	
66) 1,2-Dichlorobenzene-d4	11.564	152	167199	50.812	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	101.620%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110623\
Data File : VV032854.D
Acq On : 06 Nov 2023 15:35
Operator : SY/MD
Sample : 05259-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YCGA9

Quant Time: Nov 07 01:36:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Nov 07 01:35:20 2023
Response via : Initial Calibration

