

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110624\
 Data File : VV037932.D
 Acq On : 06 Nov 2024 16:46
 Operator : SY/MD
 Sample : P4714-13
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AH5

Quant Time: Nov 06 22:46:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 06 22:44:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	99171	5.000	ug/L	0.01
28) Chlorobenzene-d5	8.786	117	104615	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	47566	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	41185	5.426	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	= 108.600%	
7) Chloroethane-d5	1.529	69	32767	5.296	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	= 106.000%	
11) 1,1-Dichloroethene-d2	2.056	65	18553	5.632	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	= 112.600%	
20) 2-Butanone-d5	3.802	46	124590	88.063	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	= 176.120%#	
24) Chloroform-d	4.246	84	72528	5.242	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 104.800%	
26) 1,2-Dichloroethane-d4	4.947	65	39496	6.387	ug/L	0.01
Spiked Amount	5.000	Range	70 - 130	Recovery	= 127.800%	
32) Benzene-d6	4.960	84	132802	4.922	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 98.400%	
36) 1,2-Dichloropropane-d6	5.989	67	41463	5.263	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	= 105.200%	
41) Toluene-d8	7.246	98	109253	4.506	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 90.200%	
43) trans-1,3-Dichloroprop...	7.571	79	14679	4.818	ug/L	0.02
Spiked Amount	5.000	Range	55 - 130	Recovery	= 96.400%	
46) 2-Hexanone-d5	8.027	63	92952	78.516	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	= 157.040%#	
56) 1,1,2,2-Tetrachloroeth...	10.149	84	51193	8.059	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	= 161.200%#	
66) 1,2-Dichlorobenzene-d4	11.558	152	46556	5.934	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	= 118.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV110624\
Data File : VV037932.D
Acq On : 06 Nov 2024 16:46
Operator : SY/MD
Sample : P4714-13
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AH5

Quant Time: Nov 06 22:46:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 06 22:44:19 2024
Response via : Initial Calibration

