

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV110720\
 Data File : VV019270.D
 Acq On : 05 Nov 2020 11:28
 Operator : SY/MD
 Sample : VV1107WBL01
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK73

Quant Time: Nov 05 23:03:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Nov 05 22:32:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	241781	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	237957	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	118873	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	92911	48.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.72%
7) Chloroethane-d5	1.58	69	77895	52.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	104.38%
11) 1,1-Dichloroethene-d2	2.12	63	141462	38.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.40%
21) 2-Butanone-d5	3.90	46	123227	104.71	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.71%
24) Chloroform-d	4.37	84	174724	48.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.78%
26) 1,2-Dichloroethane-d4	5.05	65	125622	50.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.88%
32) Benzene-d6	5.07	84	330543	48.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.94%
36) 1,2-Dichloropropane-d6	6.09	67	106594	52.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.92%
41) Toluene-d8	7.34	98	302234	46.61	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.22%
43) trans-1,3-Dichloropropene-	7.64	79	54411	46.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.44%
47) 2-Hexanone-d5	8.11	63	78980	90.61	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.61%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	126785	46.72	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.44%
64) 1,2-Dichlorobenzene-d4	11.65	152	126519	50.66	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.32%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV110720\
Data File : VV019270.D
Acq On : 05 Nov 2020 11:28
Operator : SY/MD
Sample : VV1107WBL01
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK73

Quant Time: Nov 05 23:03:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Nov 05 22:32:19 2020
Response via : Initial Calibration

