

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV110720\
 Data File : VV019286.D
 Acq On : 05 Nov 2020 18:19
 Operator : SY/MD
 Sample : VIBLK73
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK73

Quant Time: Nov 06 00:01:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Nov 05 22:32:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	233273	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	229921	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	113090	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	87161	47.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.04%
7) Chloroethane-d5	1.58	69	74127	51.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.94%
11) 1,1-Dichloroethene-d2	2.12	63	129545	36.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.46%
21) 2-Butanone-d5	3.90	46	126790	111.67	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.67%
24) Chloroform-d	4.37	84	167096	48.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.92%
26) 1,2-Dichloroethane-d4	5.05	65	121500	50.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.12%
32) Benzene-d6	5.07	84	318293	48.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.60%
36) 1,2-Dichloropropane-d6	6.09	67	102854	52.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.78%
41) Toluene-d8	7.34	98	282845	45.14	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.28%
43) trans-1,3-Dichloropropene-	7.64	79	52090	46.29	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.58%
47) 2-Hexanone-d5	8.11	63	81865	97.20	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.20%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	130275	49.68	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.36%
64) 1,2-Dichlorobenzene-d4	11.65	152	122258	51.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.90%

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV110720\
Data File : VV019286.D
Acq On : 05 Nov 2020 18:19
Operator : SY/MD
Sample : VIBLK73
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK73

Quant Time: Nov 06 00:01:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Nov 05 22:32:19 2020
Response via : Initial Calibration

