

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110723\
 Data File : VV032861.D
 Acq On : 07 Nov 2023 10:19
 Operator : SY/MD
 Sample : VV1107MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK212

Quant Time: Nov 08 01:38:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Nov 08 01:37:19 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	373453	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	414708	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	194642	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	168147	47.338	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.680%
7) Chloroethane-d5	1.532	69	144161	48.527	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.060%
11) 1,1-Dichloroethene-d2	2.060	65	70584	48.110	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.220%
21) 2-Butanone-d5	3.783	46	139021	90.232	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.230%
24) Chloroform-d	4.249	84	274133	48.351	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.700%
26) 1,2-Dichloroethane-d4	4.941	65	174877	49.628	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.260%
32) Benzene-d6	4.960	84	536570	50.322	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.640%
36) 1,2-Dichloropropane-d6	5.989	67	175808	54.272	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.540%
41) Toluene-d8	7.243	98	522106	49.728	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.460%
43) trans-1,3-Dichloroprop...	7.551	79	80167	47.517	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.040%
47) 2-Hexanone-d5	8.024	63	98733	83.487	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.490%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	223896	45.547	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.100%
66) 1,2-Dichlorobenzene-d4	11.561	152	207270	55.259	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.520%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110723\
Data File : VV032861.D
Acq On : 07 Nov 2023 10:19
Operator : SY/MD
Sample : VV1107MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK212

Quant Time: Nov 08 01:38:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM103123WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Nov 08 01:37:19 2023
Response via : Initial Calibration

