

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV110819\
 Data File : VV013550.D
 Acq On : 08 Nov 2019 19:44
 Operator : SY/MD
 Sample : K5718-01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BE9S3

Quant Time: Nov 09 00:46:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR110819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 08 22:42:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	384045	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	383028	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	171569	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	86475	5.34	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.80%
7) Chloroethane-d5	1.58	69	79614	5.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.80%
11) 1,1-Dichloroethene-d2	2.13	63	108288	3.38	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.60%
20) 2-Butanone-d5	3.92	46	318034	54.81	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.62%
24) Chloroform-d	4.40	84	232440	5.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.80%
26) 1,2-Dichloroethane-d4	5.08	65	120866	5.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.00%
32) Benzene-d6	5.10	84	393172	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.20%
36) 1,2-Dichloropropane-d6	6.12	67	139673	5.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.80%
41) Toluene-d8	7.36	98	299730	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.40%
43) trans-1,3-Dichloropropene-	7.66	79	45607	4.53	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.60%
46) 2-Hexanone-d5	8.13	63	200640	49.78	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.56%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	109091	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	158746	5.60	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.00%

Target Compounds

					Ovalue
13) Acetone	2.19	43	5686	1.615 ug/L	79
14) Carbon disulfide	2.32	76	6632	0.096 ug/L #	80
19) 1,1-Dichloroethane	3.23	63	9966	0.186 ug/L #	94
33) Benzene	5.15	78	49341	0.415 ug/L	100
51) Chlorobenzene	8.92	112	54479	0.667 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV110819\
Data File : VV013550.D
Acq On : 08 Nov 2019 19:44
Operator : SY/MD
Sample : K5718-01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BE9S3

Quant Time: Nov 09 00:46:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR110819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 08 22:42:17 2019
Response via : Initial Calibration

