

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV110820\
 Data File : VV019290.D
 Acq On : 06 Nov 2020 11:10
 Operator : SY/MD
 Sample : VV1108MBL01
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK74

Quant Time: Nov 07 01:30:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Nov 07 01:17:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	236964	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	235277	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	113629	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	85727	45.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.04%
7) Chloroethane-d5	1.58	69	74334	50.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.62%
11) 1,1-Dichloroethene-d2	2.12	63	131208	36.63	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.26%
21) 2-Butanone-d5	3.90	46	125634	108.93	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.93%
24) Chloroform-d	4.37	84	168765	48.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.36%
26) 1,2-Dichloroethane-d4	5.05	65	121483	49.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.54%
32) Benzene-d6	5.07	84	320499	48.02	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.04%
36) 1,2-Dichloropropane-d6	6.09	67	104091	52.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.62%
41) Toluene-d8	7.33	98	286629	44.71	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.42%
43) trans-1,3-Dichloropropene-	7.64	79	52709	45.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.54%
47) 2-Hexanone-d5	8.11	63	79249	91.96	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.96%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	128300	47.82	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.64%
64) 1,2-Dichlorobenzene-d4	11.65	152	122185	51.18	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.36%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV110820\
Data File : VV019290.D
Acq On : 06 Nov 2020 11:10
Operator : SY/MD
Sample : VV1108MBL01
Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK74

Quant Time: Nov 07 01:30:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM102920WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Nov 07 01:17:34 2020
Response via : Initial Calibration

