

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110821\
 Data File : VV023274.D
 Acq On : 08 Nov 2021 11:21
 Operator : SY/MD
 Sample : VV1108WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK244

Quant Time: Nov 09 05:14:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 09 05:13:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	135307	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	133091	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	61905	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	35898	4.235	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.800%
7) Chloroethane-d5	1.565	69	31959	4.626	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.600%
11) 1,1-Dichloroethene-d2	2.105	63	53293	3.358	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.200%
20) 2-Butanone-d5	3.896	46	63604	43.554	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.100%
24) Chloroform-d	4.346	84	83052	4.598	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.000%
26) 1,2-Dichloroethane-d4	5.031	65	37677	4.638	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.800%
32) Benzene-d6	5.047	84	157606	4.615	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%
36) 1,2-Dichloropropane-d6	6.066	67	48750	4.850	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.000%
41) Toluene-d8	7.313	98	132010	4.125	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.600%
43) trans-1,3-Dichloroprop...	7.622	79	16564	4.346	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.000%
46) 2-Hexanone-d5	8.088	63	60386	43.058	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.120%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	32419	4.485	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.600%
66) 1,2-Dichlorobenzene-d4	11.625	152	51513	4.997	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110821\
Data File : VV023274.D
Acq On : 08 Nov 2021 11:21
Operator : SY/MD
Sample : VV1108WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK244

Quant Time: Nov 09 05:14:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 09 05:13:22 2021
Response via : Initial Calibration

