

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW110821\
 Data File : VW023281.D
 Acq On : 08 Nov 2021 16:10
 Operator : SY/MD
 Sample : M4445-22
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3A5

Quant Time: Nov 09 05:16:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 09 05:13:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	127453	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	128470	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	60675	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	33138	4.150	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	83.000%	
7) Chloroethane-d5	1.568	69	29635	4.554	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	91.000%	
11) 1,1-Dichloroethene-d2	2.108	63	47277	3.163	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	63.200%	
20) 2-Butanone-d5	3.889	46	80593	58.589	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	117.180%	
24) Chloroform-d	4.352	84	79902	4.696	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.000%	
26) 1,2-Dichloroethane-d4	5.037	65	37492	4.900	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.000%	
32) Benzene-d6	5.053	84	145884	4.426	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.600%	
36) 1,2-Dichloropropane-d6	6.072	67	46333	4.775	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	95.400%	
41) Toluene-d8	7.320	98	123596	4.001	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	80.000%	
43) trans-1,3-Dichloroprop...	7.625	79	15647	4.253	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	85.000%	
46) 2-Hexanone-d5	8.088	63	60662	44.811	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	89.620%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	31746	4.549	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	91.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	50121	4.961	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	99.200%	
Target Compounds						
13) Acetone	2.179	43	2590	3.081	ug/L	87
42) Toluene	7.397	91	8764	0.228	ug/L	92
53) m,p-xylene	9.149	106	1537	0.097	ug/L	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110821\
Data File : VV023281.D
Acq On : 08 Nov 2021 16:10
Operator : SY/MD
Sample : M4445-22
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3A5

Quant Time: Nov 09 05:16:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 09 05:13:22 2021
Response via : Initial Calibration

