

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW110821\
 Data File : VW023282.D
 Acq On : 08 Nov 2021 16:33
 Operator : SY/MD
 Sample : M4445-23
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3A6

Quant Time: Nov 09 05:16:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 09 05:13:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	120835	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	123652	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	60929	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	33671	4.448	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	89.000%	
7) Chloroethane-d5	1.568	69	30666	4.971	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	99.400%	
11) 1,1-Dichloroethene-d2	2.108	63	50057	3.532	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	70.600%	
20) 2-Butanone-d5	3.886	46	90149	69.125	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	138.240%#	
24) Chloroform-d	4.352	84	80428	4.985	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.800%	
26) 1,2-Dichloroethane-d4	5.034	65	38655	5.328	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	106.600%	
32) Benzene-d6	5.053	84	151186	4.765	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.400%	
36) 1,2-Dichloropropane-d6	6.072	67	48462	5.189	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	103.800%	
41) Toluene-d8	7.320	98	129232	4.347	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.000%	
43) trans-1,3-Dichloroprop...	7.625	79	16877	4.766	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	95.400%	
46) 2-Hexanone-d5	8.088	63	63860	49.011	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	98.020%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	32231	4.799	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	96.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	53952	5.318	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.400%	
Target Compounds						
					Qvalue	
13) Acetone	2.179	43	14486	18.179	ug/L	93
33) Benzene	5.104	78	13896	0.402	ug/L	100
42) Toluene	7.394	91	28439	0.769	ug/L	95
52) Ethylbenzene	9.017	91	4298	0.110	ug/L	98
53) m,p-xylene	9.140	106	5687	0.372	ug/L	98
54) o-xylene	9.548	106	5529	0.385	ug/L	94
62) 1,3,5-Trimethylbenzene	10.541	105	2078	0.072	ug/L	91
63) 1,2,4-Trimethylbenzene	10.918	105	8291	0.287	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV110821\
Data File : VV023282.D
Acq On : 08 Nov 2021 16:33
Operator : SY/MD
Sample : M4445-23
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3A6

Quant Time: Nov 09 05:16:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 09 05:13:22 2021
Response via : Initial Calibration

